



Atacama Large Millimeter Array


ACA Spectrometer Technical Specifications and Requirements

ALMA-64.00.00.00-0005-C-SPE

Version: C.1


2020-11-17

Prepared By:	Organization	Date
ACA Spectrometer Team:		
Jongsoo Kim	KASI	
Takashi Nakamoto	NAOJ	
Manabu Watanabe	NAOJ	
Approved By:	Organization	Date
Takafumi Kojima (Development Program Manager)	NAOJ	
Tetsuhiro Minamidani (EA-IET Lead)	NAOJ	
(Observatory System Engineer)	JAO	
John Carpenter (Observatory Scientist)	JAO	
Configuration Control Board Approval:	Organization	Date
(CCB Secretary)	JAO	
Released By:	Organization	Date
Alvaro Gonzalez (ALMA-EA Operations Manager)	NAOJ	

	ACA Spectrometer Technical Specifications and Requirements	Doc #: ALMA-64.00.00.00-0005-C-SPE Date: 2020-11-17 Status: Released Page: 2 of 23
--	---	---

Change Record

Version	Date	Affected Section(s)	Change Request #	Reason/Initiation/Remarks
A	2017-01-27	All	Draft A	First draft
A1	2018-11-01	All		Comprehensively revised to accommodate RIDs from PDR reviewers
A2	2018-11-16	All		NDW: various changes, formatting & corrections
A3	2018-11-21	All		Update based on telecon on 20 th / 21 st .
A4	2018-11-26	Tables 1-6 6.2 All 5.2.4		Put footnotes at Tables 1-6 Rewrite the synchronization requirement Do cosmetic format changes or minor modifications Change of sensitivity loss from 1% to 5%
A5	2019-02-22	5.2.4 Tables 1-6		Change the title of the subsection from “sensitivity loss” to “quantization loss” and quote 12% in [AD03] as the requirement of the maximum quantization loss. All the footnotes related to the quantization loss are removed.
A6	2019-03-08	5.2.4		The subsection title is changed from “quantization loss” to “sensitivity loss.” The loss is reduced from 12% to 2.1%. The verification method is changed from R to A.
A	2019-03-14	Cover page		Signature matrix edited
B	2019-05-06			Revision after release of version A
B1	2019-05-06	4.3.2 5.2.4 5.2.6 6.3.3		Insert subsection titled “Electronics environment” Make modification of the contents of the subsection “Quantization levels of input data” Insert subsection titled “Output data format” Clarify that there is no option to synthesize the frequency response from the ACA Correlator
B2	2019-06-17	5.2.3		Include the maximum output data rates
B3	2019-09-26	5.1.4		Include footnote on nutator.
B4	2019-10-09	2.3 4.4.4 4.5.2		Add [RD 09] in the list of reference documents and quote [RD 09] in 5.2.6 Put an explicit value, 20.2 kW Change of wording
B5	2019-10-15	6.2		Restate synchronization based on “metaframe bit” and [RD10] is included.
B.6	2019-11-25	2 7.4		TN-NAOJ: ICD lists are now integrated into AD list (RID-1254). Changed the title of section 7.4 (RID-1259).
B.7	2019-11-26	7.5		TN-NAOJ: Replaced “MTD” with “MDT” (RID-1258).

	ACA Spectrometer Technical Specifications and Requirements	Doc #: ALMA-64.00.00.00-0005-C-SPE Date: 2020-11-17 Status: Released Page: 3 of 23
--	---	---

B.8	2020-01-06	5.1.5 5.1.4		JK_KASI: The requirement of the spectral dynamic range is changed as suggested in RID-1291 In response to RID-444 , remove the footnote added in B.3 In response to RID-1207 , the requirement is revised. The dump time columns in Table 1, 2, 3, and 7 are removed.
B.9	2020-05-27	5.2.7		In response to RID-1204 , Jongsoo added the subsection 5.2.7 Non-linearity correction
B.10	2020-06-10			Jongsoo made a version B to be released.
C.1	2020-11-09	5.2.5		JK added sensitivity loss due to imperfect delay correction



	ACA Spectrometer Technical Specifications and Requirements	Doc #: ALMA-64.00.00.00-0005-C-SPE Date: 2020-11-17 Status: Released Page: 4 of 23
--	---	---

Table of Contents

1	INTRODUCTION	5
1.1	Purpose	5
2	RELATED DOCUMENTS AND DRAWINGS.....	5
2.1	Applicable Documents	5
2.2	Reference Documents	6
2.3	Abbreviations and Acronyms	7
2.4	Requirements Numbering.....	8
2.5	Verb Conventions	8
2.6	Order of Precedence.....	8
3	DESCRIPTION	9
4	GENERAL REQUIREMENTS	11
4.1	Operation modes	11
4.2	Safety Requirements	11
4.3	Operating Conditions	12
4.4	Electrical and Mechanical Requirements	13
4.5	Monitor and Control Requirements	14
5	FUNCTIONAL REQUIREMENTS.....	15
5.1	ALMA Overview Requirements.....	15
5.2	Requirements Specific to the ACA Spectrometer	16
6	COMPATIBILITY REQUIREMENT.....	17
6.1	Input Optical Interface.....	17
6.2	Clock and Reference Signal.....	17
6.3	Compatibilities with 64-antenna Correlator	17
7	RELIABILITY AND MAINTENANCE REQUIREMENTS	19
7.1	Line Replaceable Units (LRUs).....	19
7.2	Mean Time Between Failures (MTBF).....	19
7.3	Lifetime	19
7.4	Failure Diagnosis and Replacement.....	19
7.5	Maintenance.....	19

	ACA Spectrometer Technical Specifications and Requirements	Doc #: ALMA-64.00.00.00-0005-C-SPE Date: 2020-11-17 Status: Released Page: 5 of 23
--	---	---

1 INTRODUCTION

1.1 PURPOSE


This document summarizes the specifications and design requirements for the ACA Spectrometer. Many of the requirements are derived from ALMA project requirements and ACA Correlator requirements, and restated in this document. The rest of them are specific to the ACA Spectrometer.

2 RELATED DOCUMENTS AND DRAWINGS

2.1 APPLICABLE DOCUMENTS

The requirements in the following documents are applicable to the ACA Spectrometer unless indicated otherwise. In the case of a conflict between this document and the following documents then this one takes precedence.


No	Document Title	Reference
AD 01	ALMA System General Safety Design Specification	ALMA-10.08.00.00-003-B-SPE
AD 02	ALMA System: Electromagnetic Compatibility (EMC) Requirements	ALMA-80.05.01.00-001-B-SPE
AD 03	ALMA System Technical Requirements	ALMA-80.04.00.00-005-C-SPE
AD 04	ALMA Environmental Specification	ALMA-80.05.02.00-001-B-SPE
AD 05	ALMA System: Electrical Design Requirements	ALMA-80.05.00.00-005-C-SPE
AD 06	ALMA Power Quality (Compatibility Levels) Specifications	ALMA-80.05.00.00-001-C-SPE
AD 07	Standard for Plugs, Socket-outlets, and Couplers	ALMA-80.05.00.00-004-B-STD
AD 08	ALMA Product Assurance Requirements	ALMA-80.11.00.00-001-D-GEN
AD 09	Interface Control Document Between Back End and ACA Spectrometer	ALMA-50.00.00.00-64.00.00.00-D-ICD
AD 10	Interface Control Document Between Site (AOS Technical Building) and ACA Correlator	ALMA-20.01.02.00-62.00.00.00-C-ICD
AD 11	Interface Control Document Between ACA Correlator and ACA Spectrometer	ALMA-62.00.00.00-64.00.00.00-A-ICD

	ACA Spectrometer Technical Specifications and Requirements	Doc #: ALMA-64.00.00.00-0005-C-SPE Date: 2020-11-17 Status: Released Page: 6 of 23
--	---	---

AD 12	Interface Control Document Between ACA Spectrometer and Computing / ACA Spectrometer Control (ASC) subsystem	ALMA-64.00.00.00-70.44.00.00-B-ICD
-------	--	------------------------------------


2.2 REFERENCE DOCUMENTS

No	Document Title	Reference
RD 01	ALMA Documentation Standards	ALMA-80.02.00.00-003-G-STD
RD 02	ALMA Product Tree	ALMA-80.03.00.00-001-R-LIS
RD 03	ACA Correlator Structural Analysis Report (including Seismic Supporting Structure)	Not ALMA doc.
RD 04	ALMA Scientific Specifications and Requirements	ALMA-90.00.00.00-001-A-SPE
RD 05	ALMA System Design Description	ALMA-80.04.00.00-002-D-SPE
RD 06	Observational Modes Supported by the ALMA Correlator	ALMA memo 556
RD 07	64 Antenna Correlator Specifications and Requirements	ALMA-60.00.00.00-001-B-SPE
RD 08	Technical Specification For the Design and Manufacturing of the ALMA NUTATOR SYSTEM	ALMA-35.03.00.00-002-B-SPE
RD 09	Science Data Model Binary Data Format written by M. Pokorny and J. Pisano	Version 2.0
RD 10	ACA Spectrometer Module Timestamping Accuracy	CORL-64.00.00.00-0021-A-REP

	ACA Spectrometer Technical Specifications and Requirements	Doc #: ALMA-64.00.00.00-0005-C-SPE Date: 2020-11-17 Status: Released Page: 7 of 23
--	---	---

2.3 ABBREVIATIONS AND ACRONYMS

ACA	Atacama Compact Array
AD	Applicable Document
ALMA	Atacama Large Millimeter/submillimeter Array
AOS	Array Operation Site
ASC	ACA Spectrometer Control
ASM	ACA Spectrometer Module
BDF	Binary Data Format
COTS	Commercial Off-The-Shelf
DRX	DTS Receiver
DRXP	DRX board with PCI-express bus interface
DTS	Data Transmission Subsystem
EMC	Electro-Magnetic Compatibility
FDM	Frequency Division Mode
GPU	Graphic Processing Unit
ICD	Interface Control Document
JAO	Joint ALMA Observatory
KASI	Korea Astronomy and Space Science Institute
LRU	Line Replaceable Unit
M&C	Monitor and Control
MDT	Mean Down Time
MTBF	Mean Time Between Failures
NAOJ	National Astronomical Observatory of Japan
PCI	Peripheral Component Interconnect
RD	Reference Document
RFI	Radio Frequency Interference
RMS	Root Mean Square
TDM	Time Division Mode
TP	Total Power

	ACA Spectrometer Technical Specifications and Requirements	Doc #: ALMA-64.00.00.00-0005-C-SPE Date: 2020-11-17 Status: Released Page: 8 of 23
--	---	---

2.4 REQUIREMENTS NUMBERING

The requirements within the present document are numbered according to the following code:
[CORL-64.00.00.00-XXXXX-YY / Z]

Where:

CORL is the project code for correlator documents as defined in [RD 01];

CORL-64 represents the level 1 hierarchy for the ACA Spectrometer as defined in [RD 02]

XXXXX is a consecutive number 00010, 00020, ... (the nine intermediate numbers remaining available for future revisions of this document);

YY describes the requirement revision. It starts with 00 and is incremented by one with every requirement revision;

Z(Z1, Z2,....) describes the requirement verification method(s), where **A** stands for analysis of subsystem or product, **I** for inspection, **R** for review of design and **T** for test.

All requirements in this document trace to the Top Level requirements to pass data from the ACA Spectrometer to the ALMA computing system.


2.5 VERB CONVENTIONS

"Shall" and "must" are used when a specification or provision is mandatory. The verbs "should" and "may" indicate a specification or provision that is not mandatory. "Will" is used to express a declaration of purpose on the part of the design activity.

2.6 ORDER OF PRECEDENCE

In the event of conflict between the text of this document and applicable documents, the following order of precedence shall apply:

1. This document
2. Applicable document

	ACA Spectrometer Technical Specifications and Requirements	Doc #: ALMA-64.00.00.00-0005-C-SPE Date: 2020-11-17 Status: Released Page: 9 of 23
--	---	---

3 DESCRIPTION

The ACA Spectrometer shall provide auto-correlation and cross-antenna correlation outputs of digitized data streams from the four ACA TP antennas. It will largely replace the functionality of the ACA Correlator for the TP antennas. It should support all the observing modes provided by the 64-input Correlator and the ACA Correlator. Compared with the ACA Correlator, the ACA Spectrometer will provide a better signal-to-noise ratio of the auto-correlation outputs due to its 32-bit floating-point calculation, better flexibility to implement new capabilities for future upgrade due to its high-level software nature, and easy development and maintenance due to its use of COTS.

Figure 1 represents a high-level view of the ACA Spectrometer and its interfaces to the ALMA system [AD 09, AD 10, AD 11 & AD 12]. The red and blue boxes are the pre-existing facilities and the orange boxes are to be added. The hardware components that form part of the ACA Spectrometer are M&C computers, module servers, two network switches, and data acquisition (DRXP) cards installed in each module server. The optical signal splitters belong to the Backend. Signals from antennas go through the optical signal splitters where they are split into two paths; one path goes to the ACA Correlator and the other to the ACA Spectrometer. In this way the ACA Spectrometer can operate independently of the ACA Correlator.





Figure 1 High-level view of the ACA Spectrometer and its interfaces to ALMA. The new components to be installed for the ACA Spectrometer are indicated by orange boxes: M&C computers, module servers, optical signal splitters, and network switches. Arrows indicate network data flow (blue), optical signals (red), and analog timing signals (green). The optical signal splitters allow the ACA Spectrometer to operate independently of the ACA Correlator. The 48 ms timing signal is delivered to the DRXP cards through a daisy-chained RS-485 cable.

	ACA Spectrometer Technical Specifications and Requirements	Doc #: ALMA-64.00.00.00-0005-C-SPE Date: 2020-11-17 Status: Released Page: 11 of 23
--	---	--

4 GENERAL REQUIREMENTS

4.1 OPERATION MODES

The ACA Spectrometer shall be used with the following modes.

4.1.1 Operational and Test

[CORL-64.00.00.00-00020-00/I,R]

The Operational mode applies during normal observations with the ACA Spectrometer. In this mode, all the specifications and requirements described in this document shall apply, unless otherwise stated. Test mode is the same as Operational mode except that valid data may not be available. The Operational or Test mode is also termed “on”.

4.1.2 Non-Operational

[CORL-64.00.00.00-00030-00/I,R]

The Non-Operational mode for the ACA Spectrometer identifies the state in which the ACA is neither Operational mode nor Test mode. The Non-Operational mode is also termed “off”.

4.2 SAFETY REQUIREMENTS

[CORL-64.00.00.00-00040-00/R]

The ACA Spectrometer design shall be in compliance with ALMA System General Safety Design Specification ([AD 01]).

4.2.1 Seismic requirement


[CORL-64.00.00.00-00050-00/R]

The ACA Spectrometer system and optical signal splitters shown in Figure 1 shall be installed at the existing four computer racks in the AOS building. The racks and their support structure with the added mass shall satisfy the ALMA seismic requirements ([AD 04]). Additionally, the distribution and masses of the spectrometer hardware in the ACA correlator racks shall be consistent with the analysis in [RD 03].

4.2.2 Emergency requirement

[ALMA-64.00.00.00-00055-00/R]

A remote shutdown method of the ACA Spectrometer shall be implemented as a part of the M&C system for regular operations. It is not required to have an emergency stop button.

	ACA Spectrometer Technical Specifications and Requirements	Doc #: ALMA-64.00.00.00-0005-C-SPE Date: 2020-11-17 Status: Released Page: 12 of 23
--	---	--

4.3 OPERATING CONDITIONS

4.3.1 Start up time

[CORL-64.00.00.00-00060-00/R]

The ACA Spectrometer shall perform pre-defined startup procedure and self tests within 15 minutes after applying power. Automatic power up sequence shall be provided, if necessary, to insure restoration to normal operation without manual on-site intervention.

4.3.2 Electronics environment

[CORL-64.00.00.00-00080/R]

The design of the ACA Spectrometer electronics shall comply with the applicable electronics design requirement [AD 05].

4.3.3 Altitude

[CORL-64.00.00.00-00090-00/I,R]

The operating altitude of the ACA Spectrometer shall be 0 – 5200m.

4.3.4 Thermal environment

[CORL-64.00.00.00-00100-00/I,R]

The ACA Spectrometer shall meet all of its operational performance requirements at ambient temperatures between 10 °C and 28 °C. In Non-Operation mode, the ACA Spectrometer shall survive at ambient temperatures of -30 °C to 40 °C ([AD 04]).

4.3.5 Relative humidity


[CORL-64.00.00.00-00110-00/I,R]

The ACA Spectrometer shall meet its performance with a non-condensing relative humidity between 20% and 80%.

4.3.6 EMC

[CORL-64.00.00.00-00120-00/R,T]

Electromagnetic radiation from the ACA Spectrometer electronics shall be in compliance with project EMC and RFI standards ([AD 02]).

	ACA Spectrometer Technical Specifications and Requirements	Doc #: ALMA-64.00.00.00-0005-C-SPE Date: 2020-11-17 Status: Released Page: 13 of 23
--	---	--

4.3.7 Storage and shipping condition

For storage mode, the same specifications and requirements as for non-operational mode apply, unless otherwise stated.

[CORL-64.00.00.00-000150-00/R]

The ACA Spectrometer modules shall meet ALMA environmental requirements in the on modes (both Operational and Test modes) and shipping requirements in the Non-Operational mode ([AD 04]).

4.4 ELECTRICAL AND MECHANICAL REQUIREMENTS

The interface between the ACA Correlator and the ACA Spectrometer are described in [AD 11], and the ACA Spectrometer shall meet the following electrical and mechanical requirements.

4.4.1 Mass

[CORL-64.00.00.00-00160-00/I,R]

The mass requirement shall comply with the one described in [AD 10].

4.4.2 Physical volume

[CORL-64.00.00.00-00170-00/I]

All the hardware components except the cables of the ACA Spectrometer shall be installed in the pre-existing four computing racks in the AOS building. The rearrangement of the existing computers and auxiliary equipment in the racks are allowed ([AD 11]).


4.4.3 Electric power

[CORL-64.00.00.00-00172-00/R]

The electrical power requirements of the ACA Spectrometer shall be in accordance with [AD 10].

[CORL-64.00.00.00-00180-00/T]

The total maximum amount of electric power required for the four computing racks in the AOS building is [AD 10], which takes into account of the pre-existing computers for the ACA Correlator, rack-mountable hardware components for the ACA Spectrometer and all other equipment.

	ACA Spectrometer Technical Specifications and Requirements	Doc #: ALMA-64.00.00.00-0005-C-SPE Date: 2020-11-17 Status: Released Page: 14 of 23
--	---	--

4.4.4 Thermal load

[CORL-64.00.00.00-00190-00/I,R]

The maximum thermal load from the consumption of electric power in the computer racks in terms of calorific value is 20.2 kW [AD 10].

4.5 MONITOR AND CONTROL REQUIREMENTS


[CORL-64.00.00.00-00200-00/R]

Monitor and control capabilities at the level of LRUs shall be provided to check the status of and to diagnose the ACA Spectrometer equipment in the event of failure.

4.5.1 Overheating requirement

[CORL-64.00.00.00-00210-00/I,R]

Temperatures of critical equipment shall be sensed and reported to the M&C system. A drop-off in power of the equipment shall be handled directly or at least internally to the subsystem for the sake of damage control.

	ACA Spectrometer Technical Specifications and Requirements	Doc #: ALMA-64.00.00.00-0005-C-SPE Date: 2020-11-17 Status: Released Page: 15 of 23
--	---	--

5 FUNCTIONAL REQUIREMENTS

5.1 ALMA OVERVIEW REQUIREMENTS

5.1.1 IF bandwidth

[CORL-64.00.00.00-00220-00/R]

The input bandwidth per baseband pair (dual polarization) per antenna shall be 2 x 2 GHz. The total number of the baseband pairs per antenna and total bandwidth per antenna shall be 4 and 16 GHz, respectively.

5.1.2 Spectral resolution

[CORL-64.00.00.00-00230-00/R]

The ACA Spectrometer shall provide all the spectral resolutions offered in the observation modes in Table 1-7 at the end of this document.

5.1.3 Spectral windows

[CORL-64.00.00.00-00230-00/R]

The ACA Spectrometer (as the ACA Correlator) should support the capability to process 64 spectral windows per baseband pair for the future.

5.1.4 Minimum integration time

[CORL-64.00.00.00-00240-00/R]

The ACA Spectrometer should provide minimum integration times of 1 msec for a TDM and 32 msec for FDMs of auto-correlation, and 96 msec for a TDM of cross-correlation.


5.1.5 Spectral dynamic range

[CORL-64.00.00.00-00241-00/R]

The spectral dynamic range shall be 20000:1. Here the spectral dynamic range is defined as the ratio of the peak spectral line amplitude to the RMS noise in the spectrum.

5.1.6 Spectrometer products

[CORL-64.00.00.00-00250-00/R]

	ACA Spectrometer Technical Specifications and Requirements	Doc #: ALMA-64.00.00.00-0005-C-SPE Date: 2020-11-17 Status: Released Page: 16 of 23
--	---	--

The ACA Spectrometer shall generate products, with two digitized data streams from a baseband pair of one antenna, $X_i X_i^*$ or $Y_i Y_i^*$ for single polarization, $X_i X_i^*$ and $Y_i Y_i^*$ for dual polarization, and $X_i X_i^*$, $Y_i Y_i^*$, $X_i Y_i^*$ and $Y_i X_i^*$ for full polarization. For the calibration purpose of the TP antennas, it shall generate $X_i X_j^*$, $Y_i Y_j^*$, $X_i Y_j^*$, $Y_i X_j^*$ for only TDMs. Here, X_i and Y_j represent Fourier transformed polarized streams detected with the horizontal X direction of i-th antenna and the vertical Y direction of j-th antenna, respectively, and a sign * is a notation for the complex-conjugate operation. The i and j indices cover from 1 to 4 of the four TP antennas.

5.2 REQUIREMENTS SPECIFIC TO THE ACA SPECTROMETER

5.2.1 Number of antennas

[CORL-64.00.00.00-00260-00/R]

The ACA Spectrometer shall process the digitized IF signals from four ACA TP antennas.

5.2.2 Number of correlations

[CORL-64.00.00.00-00270-00/R]

The numbers of auto-correlations including cross polarization products per baseband pair per spectral window per antenna shall be 1 for single polarization, 2 for dual polarization, and 4 for full polarization. The number of cross-correlations per baseband pair per spectral window shall be 24 (= 6 (number of pairs from 4 antennas) x 4 (full polarization products)).

5.2.3 Quantization levels of input data

[CORL-64.00.00.00-00290-00/R]

The quantization levels of input data should be eight levels (3-bit digitization).

5.2.4 Maximum output data rates


[CORL-64.00.00.00-00300-00/R]

The TDM modes for the ACA Spectrometer with 1 m sec dump time in Table 1-3 generate a maximum output data rate from the ACA Spectrometer. The maximum output data rate shall not exceed 17 MB/sec.

5.2.5 Sensitivity Loss

[CORL-64.00.00.00-00305-00/A]

The sensitivity loss of the ACA Spectrometer should be smaller than 2.1%. (There is an additional 4% loss due to the 3-bit sampling at the antennas.)

	ACA Spectrometer Technical Specifications and Requirements	Doc #: ALMA-64.00.00.00-0005-C-SPE Date: 2020-11-17 Status: Released Page: 17 of 23
--	---	--

[CORL-64.00.00.00-00308-00/A]

The sensitivity loss of the ACA Spectrometer due to imperfect delay correction shall be smaller than 1%.

5.2.6 Output data format

[CORL-64.00.00.00-00310-00/R,T]

The ACA Spectrometer shall deliver its output data in the BDF. The details of the format are described in [RD 09].

5.2.7 Non-linearity correction

[CORL-64.00.00.00-00320-00/R,T]

A total power measurement based on 3-bit digitized signal with nonlinearity correction shall have less than 1 % error relative to the measurement of an analogue power meter over a dynamic range of 10 dB.

6 COMPATIBILITY REQUIREMENT

6.1 INPUT OPTICAL INTERFACE

[CORL-64.00.00.00-00340-00/I,T]

The ACA Spectrometer shall get digitized signals through 48 optical fiber cables from four ACA TP antennas. The detailed requirements of the input optical interface are described in [AD 09].

6.2 CLOCK AND REFERENCE SIGNAL

[CORL-64.00.00.00-00350-00/R,T]


The ACA Spectrometer shall be synchronized to the ALMA system based on a so-call “metaframe bit” in incoming DTS data streams, which is turned on 1 only for every 48 msec (see details in [RD10]).

6.3 COMPATIBILITIES WITH 64-ANTENNA CORRELATOR

6.3.1 Correlator/Spectrometer modes

[CORL-64.00.00.00-00360-00/R]

The ACA Spectrometer shall provide auto-correlation modes listed in Tables 1-6, most of which are the same as the modes provided by the 64-antenna Correlator using Tunable Filter

	ACA Spectrometer Technical Specifications and Requirements	Doc #: ALMA-64.00.00.00-0005-C-SPE Date: 2020-11-17 Status: Released Page: 18 of 23
--	---	--

Bank in terms of bandwidth and channel resolution (see Table 1-7 in [RD 06] which supersede the mode tables in [RD 07]). In order to support a top-level requirement of 1 msec output rate of auto-correlation ([AD 03]), the lowest resolution modes in Tables 1-3 support a minimum dump time of 1 msec. The ACA Spectrometer shall provide cross-correlation modes listed in Table 7 for calibration purpose, which are newly added modes.

6.3.2 Channel-averaging factors


[CORL-64.00.00.00-00370-00/R]

The ACA Spectrometer shall provide the capability for the channel-averaging (also known as spectral averaging) with binning factors = 2, 4, 8, 16, 32, 64.

6.3.3 Frequency profile synthesis

[CORL-64.00.00.00-00370-00/A,R]

The ACA Spectrometer shall be implemented with an option to synthesize the frequency response profile of the 64-input Correlator with a window function. However, there is no need to synthesize the frequency response profile of the ACA Corrector, because the frequency responses from the ACA Correlator and Spectrometer are basically same. It shall also generate its own frequency response profile.

	ACA Spectrometer Technical Specifications and Requirements	Doc #: ALMA-64.00.00.00-0005-C-SPE Date: 2020-11-17 Status: Released Page: 19 of 23
--	---	--

7 RELIABILITY AND MAINTENANCE REQUIREMENTS

7.1 LINE REPLACEABLE UNITS (LRUs)

[CORL-64.00.00.00-00390-00/I]

LRUs shall be used as far as possible and will include the ASC, ASM, power suppliers and cooling fans of the ASC and ASM, GPU cards, and DRXP cards. All of them are COTS except the DRXPs.

7.2 MEAN TIME BETWEEN FAILURES (MTBF)

[CORL-64.00.00.00-00400-00/A]

The top level requirement of the availability of the Array in [AD 03] shall be larger than 85% with a goal of 95% for the steady ALMA operations. Assuming that 1/3 of the TP Array down time is due to the ACA Spectrometer and half of the down time is due to the hardware failures of the ACA Spectrometer, the total time to repair the failed hardware in a year is $\frac{1}{3} * \frac{1}{2} * 0.05 * 24 \text{ hours} * 365 \text{ days/year} = 73 \text{ hours / year}$.

Consequently, and assuming that the average time to execute a hardware repair is 2 hours + 12 hours = 14 hours, a minimum MTBF for the ACA Spectrometer against hardware failure shall be $14 \text{ hours} / (73 \text{ hours/year}) = 0.2 \text{ year}$.

7.3 LIFETIME

[CORL-64.00.00.00-00410-00/A]

The lifetime of the ACA Spectrometer shall be more than 15 years.

7.4 FAILURE DIAGNOSIS AND REPLACEMENT

[CORL-64.00.00.00-00420-00/A,R]

The failure diagnosis and software maintenance shall be done remotely. The failure diagnosis program shall not exceed one hour. The replacement of LRUs should be done with two persons for one hour. The adequate preparation of spare LRUs shall be provided to meet the lifetime of the ACA Spectrometer.

7.5 MAINTENANCE

[CORL-64.00.00.00-00430-00/A,R]

The ACA Spectrometer should need the overall maintenance once a year during the ALMA maintenance period, and the MDT shall not exceed 1 day. The ACA Spectrometer should have simple regular maintenance once a month, and the MDT of the monthly maintenance shall not exceed 1 hour.



**ACA Spectrometer
Technical Specifications and Requirements**

Doc #: ALMA-64.00.00.00-0005-C-SPE
Date: 2020-11-17
Status: Released
Page: 20 of 23

Table 1*. Auto-correlation mode chart with one baseband pair for single polarization

Mode #	Number of sub-channel filters	Total BW [MHz]	Number of Spectral Points	Spectral Resolution [kHz]	Velocity Resolution at 230 GHz [km/s]
1	32	2000	8192	244	0.32
19	32	2000	4096	488	0.64
38	32	2000	2048	976	1.28
2	16	1000	8192	122	0.16
20	16	1000	4096	244	0.32
39	16	1000	2048	488	0.64
53	16	1000	1024	976	1.28
3	8	500	8192	61	0.08
21	8	500	4096	122	0.16
40	8	500	2048	244	0.32
54	8	500	1024	488	0.64
4	4	250	8192	30	0.04
22	4	250	4096	61	0.08
41	4	250	2048	122	0.16
55	4	250	1024	244	0.32
5	2	125	8192	15	0.02
23	2	125	4096	30	0.04
42	2	125	2048	61	0.08
56	2	125	1024	122	0.16
6	1	62.5	8192	7.6	0.01
24	1	62.5	4096	15	0.02
43	1	62.5	2048	30	0.04
57	1	62.5	1024	61	0.08
25	1	31.25	8192	3.8	0.005
58	1	31.25	2048	15	0.02
68	TDM	2000	256	7813	10.2

*All the observing modes supported by the ACA Spectrometer have data streams sampled with the Nyquist time interval.



**ACA Spectrometer
Technical Specifications and Requirements**

Doc #: ALMA-64.00.00.00-0005-C-SPE
Date: 2020-11-17
Status: Released
Page: 21 of 23

Table 2*. Auto-correlation mode chart with one baseband pair for dual polarization

Mode #	Number of sub-channel filters	Total BW [MHz]	Number of Spectral Points	Spectral Resolution [kHz]	Velocity Resolution at 230 GHz [km/s]
7	32	2000	4096	488	0.64
8	16	1000	4096	244	0.32
26	16	1000	2048	488	0.64
44	16	1000	1024	976	1.28
9	8	500	4096	122	0.16
27	8	500	2048	244	0.32
45	8	500	1024	488	0.64
59	8	500	512	976	1.28
10	4	250	4096	61	0.08
28	4	250	2048	122	0.16
46	4	250	1024	244	0.32
60	4	250	512	488	0.64
11	2	125	4096	30	0.04
29	2	125	2048	61	0.08
47	2	125	1024	122	0.16
61	2	125	512	244	0.32
12	1	62.5	4096	15	0.02
30	1	62.5	2048	30	0.04
48	1	62.5	1024	61	0.08
62	1	62.5	512	122	0.16
31	1	31.25	4096	7.6	0.01
63	1	31.25	1024	30	0.04
69	TDM	2000	128	15600	20.4

* All the observing modes supported by the ACA Spectrometer have data streams sampled with the Nyquist time interval.



**ACA Spectrometer
Technical Specifications and Requirements**

Doc #: ALMA-64.00.00.00-0005-C-SPE
Date: 2020-11-17
Status: Released
Page: 22 of 23

Table 3*. Auto-correlation mode chart with one baseband pair for full polarization

Mode #	Number of sub-channel filters	Total BW [MHz]	Number of Spectral Points	Spectral Resolution [kHz]	Velocity Resolution at 230 GHz [km/s]
13	32	2000	2048	976	1.28
14	16	1000	2048	488	0.64
32	16	1000	1024	976	1.28
15	8	500	2048	244	0.32
33	8	500	1024	488	0.64
16	4	250	2048	122	0.16
34	4	250	1024	244	0.32
17	2	125	2048	61	0.08
35	2	125	1024	122	0.16
51	2	125	512	244	0.32
18	1	62.5	2048	30	0.04
36	1	62.5	1024	61	0.08
52	1	62.5	512	122	0.16
66	1	62.5	256	244	0.32
37	1	31.25	2048	15	0.02
67	1	31.25	512	61	0.08
70	TDM	2000	64	31250	40.8

* All the observing modes supported by the ACA Spectrometer have data streams sampled with the Nyquist time interval.


	ACA Spectrometer Technical Specifications and Requirements	Doc #: ALMA-64.00.00.00-0005-C-SPE Date: 2020-11-17 Status: Released Page: 23 of 23
--	---	--

Table 4*. Multi-resolution mode chart with one baseband pair for single polarization

Mode #	Minimum size of correlation	Number of sub-channel filters	Bandwidth [MHz]	Number of Spectral Points
2	1/2	16	1000	8192
3	1/4	8	500	8192
4	1/8	4	250	8192
5	1/16	2	125	8192
6	1/32	1	62.5	8192
25	1/32	1	31.25	8192

* All the observing modes supported by the ACA Spectrometer have data streams sampled with the Nyquist time interval.

Table 5*. Multi-resolution mode chart with one baseband pair for dual polarization

Mode #	Minimum size of correlation	Number of sub-channel filters	Bandwidth [MHz]	Number of Spectral Points
9	1/2	8	500	4096
10	1/4	4	250	4096
11	1/8	2	125	4096
12	1/16	1	62.5	4096
31	1/16	1	31.25	4096

* All the observing modes supported by the ACA Spectrometer have data streams sampled with the Nyquist time interval.

Table 6*. Multi-resolution mode chart with one baseband pair for full polarization

Mode #	Minimum size of correlation	Number of sub-channel filters	Bandwidth [MHz]	Number of Spectral Points
16	1/2	4	250	2048
17	1/4	2	125	2048
18	1/8	1	62.5	2048
37	1/8	1	31.25	2048

* All the observing modes supported by the ACA Spectrometer have data streams sampled with the Nyquist time interval.

Table 7. Cross-correlation mode chart with one baseband pair for dual and full polarizations. These modes also provide auto-correlation results.

Mode #	Spectral modes	Total Bandwidth [MHz]	Correlation products	Number of Spectral Points	Minimum dump time [msec]
81*	TDM	2000	XX*, YY*	128	96
82*	TDM	2000	XX*, YY*, XY*, YX*	64	96

*Newly added modes. There are no corresponding modes in the mode charts of the 64-input Correlator.